

**टाटा मूलभूत अनुसंधान संस्थान हैदराबाद**  
**TATA INSTITUTE OF FUNDAMENTAL RESEARCH HYDERABAD**  
प्लॉट नं. ३६/पी, गोपनपल्ली ग्राम, सेरिलिंगमपल्ली मंडल, रंगा रेड्डी जिला,  
हैदराबाद – ५०० १०७, तेलंगाना.  
Plot No.36/P, Gopanpally Village, Serilingampally Mandal, Ranga Reddy District,  
Hyderabad - 500 107, Telangana, India.

टेलीफोन /Telephone: +91-40-2020 3020  
वेबसाइट /Website: [www.tifrh.res.in](http://www.tifrh.res.in)

ई-मेल /Email: [purchasegroup@tifrh.res.in](mailto:purchasegroup@tifrh.res.in)  
तिथि /Date: 21-08-2024

**Global Tender Ref. No. TIFR/PD/GT/CA24-03/24-25/M/03**

**Sub: CORRIGENDUM NO. 1, against Tender Ref. No. TIFR/PD/GT/CA24-03/24-25/M/03 for Supply, Installation and Commissioning of Electron Microscope for TIFR Hyderabad.**

Dear Bidders/Vendors,

Please refer the subject tender published on 26-07-2024, the following corrigendum to the subject is being issued:

S.No	Change from	To be change/amended as
1.	<b>Annexure-A</b> <b>(Page 3 of tender document)</b>  <b>General system requirements</b>	<b>Include</b> <ul style="list-style-type: none"><li>All the resolutions for the electron microscope, STEM detector, and FIB must be provided by specifying the methodology used for obtaining the resolution, for example using average / best-value criteria, ISO DTS 24597, 2010, derivative method. Wherever required, include the resolution under each criteria/method used.</li></ul>
2.	<b>Refer Page 4 of tender document</b>  1.6 Detector: 6. Motorized retractable STEM detector with BF,DF,HAADF (system should support mounting the prepared lamella to STEM holder through Nano manipulator without venting main chamber). Resolution of STEM detector should be atleast 0.6nm @ 30kV	1.6 Detector: 6. Motorized retractable STEM detector with BF,DF,HAADF (system should support mounting the prepared lamella to STEM holder through Nano manipulator without venting main chamber). Resolution of STEM detector should be atleast <b>0.7nm @ 30kV</b>

3.	<b>Refer Page 4 of tender document</b> 1.6 Detector: 7. Dedicated ion column detector	1.6 Detector: 7. Ion column detector
4.	Last Date for Submission of Bid: <b>27-08-2024 upto 13.00 Hrs.</b>	Last Date for Submission of Bid <b>11-09-2024 at 13.00 Hrs.</b>
5.	Date of Opening Bids (Part A): <b>27-08-2024 at 15.00 Hrs.</b>	Date of Opening Bids (Part A) <b>11-09-2024 at 15.00 Hrs.</b>

**Note: Remaining technical specifications, terms & conditions as per tender only.**

All other terms & conditions of subject tender shall remain unchanged. This Corrigendum No. 01 is an integral part of the subject tender and a copy of the same must be submitted along with the offer duly signed and stamped.

**प्रशासनक अवधकारी / ADMINISTRATIVE OFFICER  
(क्रय अनुभाग) / (PURCHASE SECTION)  
टीआईएफआर हैदराबाद / TIFR HYDERABAD**