# टाटा मूलभूत अनुसंधान संस्थान हैदराबाद TATA INSTITUTE OF FUNDAMENTAL RESEARCH HYDERABAD

प्लाट नं. ३६/पी, गोपनपल्ली ग्राम, सेररललंगमपल्ली मंडल, रं गा रे ड्डी लिला,हैदराबाद – ५०० १०७, तेलंगाना.

Plot No.36/P, Gopanpally Village, Serilingampally Mandal, Ranga Reddy District, Hyderabad - 500 107, Telangana, India.

टेलीफोन /Telephone: +91-40-2020 3020

वेबसाइट /Website: www.tifrh.res.in

ई-मेल /Email: purchasegroup@tifrh.res.in

तिति /Date: 26-09-2024

## Global Tender Ref. No. TIFR/PD/GT/IC24-04/240695

Sub: CORRIGENDUM NO. 1, against Tender Ref. No. TIFR/PD/GT/IC24-04/240695 fr "Supply, Installation and commissioning of Triple Quadrupole (QQQ) Liquid Chromatography Mass Spectrometry for TIFR Hyderabad".

Dear Bidders/Vendors,

Please refer the subject tender published on 14-09-2024, the following corrigendum to the subject is being issued:

### Page 3: Pump

- Operating pressure should be 15000 psi or better
- UV detector should be quoted with wavelength range 190 600 nm

#### Page 4: Mass Spectrometer

- Scan speed: at least 15,000 Da/sec or better
- Sensitivity: 1pg of Reserpine should have signal to noise of 5,00,000:1 and on column instrument detection limit of less than 4fg, to be present in official datasheet; or 1pg of Chloramphenicol should have signal to noise of 5,00,000:1 and on column instrument detection limit of less than 4fg, to be present in official datasheet. It should be demonstrated in front of committee and mentioned value should be there in the submitted official datasheet and should be reproduced after installation.

## Page 5: Software and data analysis tools

 Software for lipidomics and metabolomics applications as well as other related and relevant applications mentioned, that can perform quantitative analysis with statistical tests, should be provided and quoted. The software should have full-fledged capability to calculate system

Page 1 of 2

suitability parameters, both for MS/MS and UHPLC. The software should allow the users to load, process and view results from their sample datasets.

New specification to be added:

- Column oven should be able to maintain temperature between 8 – 75°C or better.

Note: Remaining technical specifications, terms & conditions as per tender only.

All other terms & conditions of subject tender shall remain unchanged. This Corrigendum No. 01 is an integral part of the subject tender and a copy of the same must be submitted along with the offer duly signed and stamped.

प्रशासवनक अवधकारी / ADMINISTRATIVE OFFICER (क्रय अनुभाग) / (PURCHASE SECTION) टीआईएफआर हैदराबाद / TIFR HYDERABAD